



QPS Evaluation Services Inc
Testing, Certification and Field Evaluation Body
Accredited in Canada, the USA, and Internationally

File
LR1638-1R1

CERTIFICATE OF COMPLIANCE
(ISO TYPE 3 CERTIFICATION SYSTEM)

Issued to	Transus Instruments B.V.
Address	Bloesemlaan 4 3897LN, Zeewolde The Netherlands
Project Number	LR1638-1R1
Product	Ultrasonic Flowmeter
Model Number	UIM Series Flowmeter / See ANNEX below for nomenclature
Electrical Ratings	See Control Drawing 06_0010
Markings	Class I Division 1 GR ABCD T4-T1 Ex ia IIC T4-T1 Ga Class I Zone 0, Ex ia IIC T4-T1 Ga -40°C ≤ Tamb ≤ +60°C -40°C ≤ Tamb ≤ +445°C see ANNEX below for derating of T-class
Applicable Standards	CSA 60079-0:19 CSA 60079-11:14 UL 60079-0 7th ed. UL 60079-11 6th ed
Factory/Manufacturing Location	Same as Applicant
Conditions of Certification	N/A

Statement of Compliance: The product(s)/equipment identified in this Certificate and described in the Certification Report covered under the above referenced project number have been investigated and found to be in compliance with the relevant requirements of the above referenced standard(s). As such, they are eligible to bear the QPS Certification Mark shown below, in accordance with the provisions of QPS's Service Agreement.

IMPORTANT NOTE: In order to maintain the integrity of the QPS Mark(s), certification will be revoked if:

- (1) Compliance to the above-mentioned Standard(s), or those identified in future QPS Standard Update Notice – SUN (QSD 55) is not maintained, or,
- (2) If the product/equipment is modified after certification is granted without prior written consent by QPS.



Issued By: Marco Erdhuizen,
 Certification Manager, Hazardous Locations [Ex Equipment] Department

Marco Erdhuizen

Signature:

Date: October 17, 2023



The SCC and IAS Accreditation Symbols are official symbols of the respective accreditation bodies, used under license.
 81 Kelfield St., Unit 8, Toronto, ON M9W 5A3 Tel: 416-241-8857; Fax: 416-241-0682
 www.qps.ca





QPS Evaluation Services Inc
 Testing, Certification and Field Evaluation Body
 Accredited in Canada, the USA, and Internationally

File
LR1638-1R1

ANNEX

Nomenclature:

UIM Electronics assembly model number		
UIME-AB-C-DEFGH		
A	x	Number of paths (1 to 4)
B	x	Application type
C	x	Meter size
D - SLOT 1	0	Not installed
	1	RS485 Option board (01-0020)
	2	RS485 IO 4..20mA option board (01-0202)
	3	Dual RS485 IO option board (01-0251)
E - SLOT 2	0	Not installed
	1	P/T option board (01-0022)
	2	4..20mA HART option board (01-0203)
	3	Dual RS485 IO option board (01-0251)
F - LCD	0	Not installed
	1	Installed
	2	SS316 enclosure with display/keypad
G	1	M20 cable gland entries
	2	1/2" NPT cable gland entries
H	x	Options, not affecting explosion safety

The maximum process temperature for each temperature class shall be limited as follows:

Temperature Class	Maximum Process Temperature
T1	445°C
T2	295°C
T3	195°C
T4	130°C

